


TECHNICAL SHEET

REF BCV02

Luer lock male-female check-valve


Date :
16/04/2025

[illegible]

Unit Label artwork code : ULBCV02



NA

$$P_{\text{bar}} \leq 3$$


NA



NA



100 / 1000

(39 x 24 x 14 cm.)



VOL

0,01

NA : Non applicable

UDI - GTIN Numbers :

Product unit : 06944262924077

Product Box: 26944262924071

Origin : P.R.C.

Shelf Life (Years) 5

Manufacturing Clean

Room class : ISO 8

Sterilization : Ethylene Oxide

In conformity with the Medical Regulation 2017/745 - Annex XI Part A - Rule 2 - CE Certificate valid up to December 2027

MDR Classification and Medical Device group
(EMDN code): I sterile-A070501-CAPS OR OBTURATORS, NON-
PERFORABLE

EC Notified Body:

0123 - TÜV SÜD Product Service GmbH - Allemagne/Germany

Manufacturer: Wenzhou KLF Medical Plastics Co., Ltd.
SRN CN-MF-000011214

Quality Management Systems Certification : ISO13485

Production and testing ISO8536-10, ISO14971, ISO 594-1, ISO 594-2,
Regulations : ISO1707, ISO11135, ISO11737-1, EN1174-1